
# 105SDG n-BUTYL AMINE



#### 1. PERFORMANCE

1) Measuring range  $\begin{array}{c} \text{1-20 ppm} \\ \text{Number of pump strokes} \end{array}$ 

2) Sampling time : 1 minute/1 pump stroke

3) Detectable limit : -

4) Shelf life : 3 years 5) Operating temperature :  $15 \sim 25 \,^{\circ}\text{C}$ 

6) Reading : Graduations printed on the tube are calibrated by Ammonia at 1 pump stroke

and n-Butyl amine concentration is determined by using a conversion chart at

1 pump stroke.

7) Colour change : Pale purple → Pale Yellow

#### 2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

## 3. CHEMICAL REACTION

By reacting with Phosphoric acid, PH indicator is discoloured. C<sub>4</sub>H<sub>9</sub>NH<sub>2</sub> + H<sub>3</sub>PO<sub>4</sub> → (R<sub>2</sub>NH<sub>3</sub>)<sub>3</sub>PO<sub>4</sub>

## 4. CALIBRATION OF THE TUBE

PERMEATION TUBE METHOD

## 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Amines	Similar stain is produced.	Higher readings are given.

