

1. PERFORMANCE

1) Measuring range 0.02-2.0% 0.01-0.9%Number of pump strokes 0.02-2.0% 0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%0.01-0.9%

3) Detectable limit : 10 ppm4) Shelf life : 3 years5) Operating temperature $: 0 \sim 40 \,^{\circ}\text{C}$

6) Reading : Graduations printed on the tube are calibrated by Ethylene oxide at 1 pump stroke

and Furan concentration is determined by using a conversion chart.

7) Colour change : Orange→Black

2. RELATIVE STANDARD DEVIATION

RSD-low: 15 % RSD-mid.: 10 % RSD-high: 5 %

3. CHEMICAL REACTION

Chromium oxide is reduced.

$$\boxed{\bigcirc \bigcirc} + Cr^{6+} + H_2SO_4 \rightarrow Cr^{3+}$$

4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

FIG.1 Influence of Toluene

5. INTERFERENCE AND CROSS SENSITIVITY

Substance		Interference	ppm	Coexistence
Aromatic hydrocarbons	FIG.1	Similar stain is produced.		Higher readings are given.
Esters	FIG.2	"		"
Ketones		"		
Alcohols	FIG.3	"		"
Halogenated hydrocarbons		Whole reagent is changed to Pale brown.	0.5%	"

(NOTE)

In case of 1/2 pump strokes, following conversion scale is available for the actual concentration.




FIG.3 Influence of Menthanol

FIG.2 Influence of Ethyl acetate