
# **DIISOBUTYL KETONE**



#### 1. PERFORMANCE

1) Measuring range 20-1,000 ppmNumber of pump strokes  $1(100\text{m}\ell)$ 

2) Sampling time : 1.5 minutes/1 pump stroke

3) Detectable limit : -

4) Shelf life : 2 years 5) Operating temperature :  $15 \sim 25 \,^{\circ}\text{C}$ 

6) Reading : Graduations printed on the tube are calibrated by Metyl ethyl ketone at 1 pump

stroke and Diisobutyl ketone concentration is determined by using a conversion

chart at 1 pump stroke.

7) Colour change : Yellow→Pale blue

#### 2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 10% RSD-high: 10%

### 3. CHEMICAL REACTION

Chromium oxide is reduced.

 $((CH_3)_2CHCH_2)_2CO + Cr^{6+} + H_2SO_4 \rightarrow Cr^{3+}$ 

## 4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

### 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Alcohol	Similar or Brown stain is produced.	Higher readings are given.
Esters	"	"
Ketones	"	"
Aromatic hydrocarbons	"	"
Halogenated hydrocarbons		Whole reagent is changed to Brown, but if the maximum end point of the Pale blue stain is discernable, the accuracy of reading is not affected.
Aliphatic hydrocarbons		"



No. 139U tube reading (ppm)