

Each detector tube is formulated with high purity reagents that absorb and react with the target gas or vapour being measured. A colorimetric stain is created that is proportional in length to the concentration. For most tubes, the concentration is read directly off the measurement scale on each tube.

