


Tube No.
900NHH

AMMONIA



1. PERFORMANCE

- 1) Measuring range : 10-80 $\mu\text{g}/\text{m}^3$
- 2) Sampling time : 24000m ℓ (400m ℓ /min \times 60min)
- 3) Detectable limit : —
- 4) Shelf life : 2 years
- 5) Operating temperature : 10 ~ 40 $^{\circ}\text{C}$
- 6) Reading : Direct reading from the scale calibrated at the sampling of 400m ℓ \times 60min
- 7) Colour change : Pale purple \rightarrow Pale yellow

2. RELATIVE STANDARD DEVIATION

RSD-low : 5 % RSD-mid. : 5 % RSD-high : 10 %

3. CHEMICAL REACTION

Indicator is discoloured by neutralization of Ammonia with acid.

4. CALIBRATION OF THE TUBE

PERMEATION TUBE METHOD

5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Amines	Similar stain is produced.	Higher readings are given.

(NOTE)

Air sampler is required for this tube.